



These products are sold for Research Use Only; not for use in diagnostic procedures

Workflow Summary

- 01 Allow the Master Mix tubes to thaw; then gently vortex to mix.
- 02 In a containment hood or dead air box, pipette 45 µL of Master Mix into individual wells of a PCR plate (use a different indexed Master Mix for each sample and control).
- 03 Add 0.2 µL EagleTaq™ DNA polymerase to each Master Mix.
- 04 Add 5 µL of DNA (at a minimum of 10 ng/µL) from the unknown samples and controls to wells containing the respective Master Mix reactions then pipette up and down 5-10 times to mix.
- 05 Add 5 µL of molecular biology grade water to the well containing the Master Mix for the no template control then pipette up and down 5-10 times to mix.
- 06 Seal the plate and amplify target DNA using the following thermal cycler program:

| Step | Temperature | Time | Cycle |
|------|-------------|------------|-------|
| 1 | 95 °C | 7 minutes | 1 |
| 2 | 95 °C | 45 seconds | 29x* |
| 3 | 60 °C | 45 seconds | |
| 4 | 72 °C | 90 seconds | |
| 5 | 72 °C | 10 minutes | 1 |
| 6 | 15 °C | ∞ | 1 |
- 07 Purify the PCR products using the Agencourt® AMPure® XP PCR Purification system. When using the LymphoTrack TRB Assay add 35 µL of particles to each 50 µL reaction, for all other LymphoTrack Assays add 50 µL of particles to each 50 µL reaction; elute DNA in 25 µL of eluant.
- 08 Quantify amplicons using the KAPA™ library quantification kit according to the kit instructions. Dilute amplicons 1:4,000 before proceeding to qPCR.
- 09 Pool equal amounts of amplicons from samples and positive and negative controls (do not include the no template control), dilute 1:1,000 then quantify the library using the KAPA library quantification kit.
- 10 Denature and dilute the library.
- 11 Load 600 µL of denatured and diluted library on the MiSeq® Reagent Cartridge.
- 12 Set up a MiSeq® sample sheet using the Illumina® Experiment Manager or provided Sample Sheet *.csv file.
- 13 Start the MiSeq® run.
- 14 Analyze and visualize the acquired data using the LymphoTrack Software for the MiSeq®.

*LymphoTrack IGHV Leader Somatic Hypermutation Assay (32 cycles).

CATALOG

7-121-0129
7-121-0139
7-121-0009
7-121-0039
7-121-0149
7-121-0089
7-121-0099
7-121-0109
7-121-0119
7-121-0059
7-121-0069
7-122-0009
7-122-0019
7-225-0009
7-225-0019
7-227-0019
7-227-0009
7-500-0009
7-500-0008

PRODUCTS

LymphoTrack® IGH FR1/2/3 Assay Kit A - MiSeq®
LymphoTrack® IGH FR1/2/3 Assay Panel - MiSeq®
LymphoTrack® IGH FR1 Assay Kit A - MiSeq®
LymphoTrack® IGH FR1 Assay Panel - MiSeq®
LymphoTrack® IGH FR1 Assay Panel B - MiSeq®
LymphoTrack® IGH FR2 Assay Kit A - MiSeq®
LymphoTrack® IGH FR2 Assay Panel - MiSeq®
LymphoTrack® IGH FR3 Assay Kit A - MiSeq®
LymphoTrack® IGH FR3 Assay Panel - MiSeq®
LymphoTrack® IGHV Leader Somatic Hypermutation Assay Kit A - MiSeq®
LymphoTrack® IGHV Leader Somatic Hypermutation Assay Panel - MiSeq®
LymphoTrack® IGK Assay Kit A - MiSeq®
LymphoTrack® IGK Assay Panel - MiSeq®
LymphoTrack® TRB Assay Kit A - MiSeq®
LymphoTrack® TRB Assay Panel - MiSeq®
LymphoTrack® TRG Assay Kit A - MiSeq®
LymphoTrack® TRG Assay Panel - MiSeq®
LymphoTrack® Software - MiSeq®
LymphoTrack® MRD Software

QUANTITY

indices 1-8 (5 sequencing reactions each)
indices 1-24 (5 sequencing reactions each)
indices 1-8 (5 sequencing reactions each)
indices 1-24 (5 sequencing reactions each)
indices 25-48 (5 sequencing reaction each)
indices 1-8 (5 sequencing reactions each)
indices 1-24 (5 sequencing reactions each)
indices 1-8 (5 sequencing reactions each)
indices 1-24 (5 sequencing reactions each)
indices 1-8 (5 sequencing reactions each)
indices 1-24 (5 sequencing reactions each)
indices 1-8 (5 sequencing reactions each)
indices 1-24 (5 sequencing reactions each)
indices 1-8 (5 sequencing reactions each)
indices 1-24 (5 sequencing reactions each)
1 CD complimentary with purchase
1 CD complimentary with purchase

Storage Conditions: -85 °C to -65 °C (DNA controls may be separated from kits and stored at 2 °C to 8 °C)

These products are sold for Research Use Only; not for use in diagnostic procedures.

©2021 Invivoscribe, Inc. (Invivoscribe®). All rights reserved. The trademarks mentioned herein are the property of Invivoscribe and/or its affiliates, or (as to the trademarks of others used herein) their respective owners. ILLUMINA® and MiSeq® are registered trademarks of Illumina, Inc. AGENCOURT® and AMPURE® are registered trademarks of Beckman Coulter, Inc. EAGLETAQ™ is a trademark of Roche. KAPA™ is a trademark of Kapa Biosystems.